

Certificate of Compliance

Certificate:

1482880 (LR 90397-1)

Master Contract:

181472

Project:

1550252

Date Issued:

2004/06/01

Issued to:

Crowcon Detection Instruments

Limited

2 Blacklands Way Abingdon Business Park

Abingdon, Oxfordshire 0X14 1DY

United Kingdom

Attention: Mr. Adrian Beasley

The products listed below are eligible to bear the CSA Mark shown



Issued by:

Authorized by: Nick Alfano, Operations

Manager

PRODUCTS

CLASS 4828 01

- SIGNAL APPLIANCES - - Combustible Gas Detection Instruments - For Hazardous

Triple Plus, Triple Plus+ and Triple Plus DLG, portable, four channel, multi-gas detector (combustibles, toxics and oxygen), battery operated (CROWCON battery pack, P/N C01253); intrinsically safe with or without integral sampling pump, Temp Code T4.

APPLICABLE REQUIREMENTS

CSA Std C22.2 No. 30-M1986 - Explosion-Proof Enclosures for Use in Class I Hazardous Locations

DQD 507 2003/01/31