



Certificate of Compliance

Certificate Number: LR 90397-6

Revision:

Date Issued: April 3, 1995

Issued To: CROWCON DETECTION INSTRUMENTS LTD.
2 Blacklands Way
Abingdon Business Park
Abingdon, Oxon
OX14 1DY, England

The products listed below are eligible to bear the CSA Mark.

Issued By: Andrew Redeker, C.E.T.
Edmonton, Alberta Canada

Signature P Andrew Redeker

CLASS

CLASS 4828 01 - SIGNAL APPLIANCES - Combustible Gas Detection Instruments - For Hazardous Locations

PRODUCTS

Class I, Groups A, B, C and D:

"DETECTIVE", portable, four channel, multi-gas detectors (combustibles, toxics and oxygen), battery operated ("CROWCON" battery pack Cat No CO1-406); Intrinsically safe, Temp Code T3A, Encl Type 3R.