

CERTIFICATE OF TEST

DATE OF ISSUE: **25 July 2011**

CERTIFICATE NUMBER: **U1780**



Environmental Test Services Ltd, Ladbroke Close, Woodley, Berks, RG5 4DX

ENVIRONMENTAL TEST SERVICES

Tel: 0118 927 2220 Fax: 0118 927 2120 E-Mail: ets@etsonline.co.uk

PAGE **1** OF **1** PAGES

APPROVED SIGNATORY

J A Fennell

Laboratory Manager

Registered in England No. 3110350

Customer: Crowcon Detection Instruments Ltd
2 Blacklands Way
Abingdon Business Park
Abingdon
Oxon OX14 1DY

Order No: PO-158364

Description of Equipment Tested:

1 off Gaspro Gas Detector

Serial No: RC2255/00/01*

**Number allocated by ETS in the absence of any other unique identification*

Tested to:

Dust Ingress Test

BS EN 60529:1991, Paragraph 13.4, IP6X Dust Tight

Results:

Post IP6X test inspection
No dust ingress was found in the enclosure

Comments:

The test item was received on 30 March 2011.
The test commenced on 31 March 2011 and completed on 1 April 2011.
The test was performed in accordance with the test specification.
ETS fitted a vacuum line to the customer fitted port on the test item.
The test duration was 8 hours.
The post test inspection was performed by ETS and witnessed by a customer representative.
A single page attachment to this certificate "DUST INGRESS TEST – IP6X: PRE AND POST TEST PHOTOGRAPHS" was issued to the customer.
The formal inspection and operational performance testing was the customer's responsibility.
All test records for the work detailed above are filed at ETS under Route Card No: 2255.

Prepared by:

N Bond

PP

Disclaimer(s):

The results recorded within this certificate apply only to the items/equipment referenced herein.

ETSGF15303 Issue 5 02/10